

# **Der's Guide**

# Shop online at omega.com

e-mail: info@omega.com For latest product manuals: www.omegamanual.info

**OM-90 SERIES** Portable Temperature and Humidity Data Loggers





# User Guide

# Contents

1.	Doc	cument Overview
2.	Pro	duct Overview4
3.	Тес	hnical Specifications6
4.	But	tons and LED Indicators7
5.	Pac	kaging List8
6.	Inst	tallation of the Software9
7.	Usir	ng the Software10
7	.1.	Front Page 10
7	.2.	Main Page
7	.3.	Configuration of a Data Logger (Start on Disconnection)
7	.4.	Saving a Configuration File14
7	.5.	Loading a Configuration from File15
7	.6.	Retrieving Logged Data16
7	.7.	Viewing Logged Data from File
7	.8.	Configuration of a Data Logger (Start at Specified Time)
7	.9.	Configuration of a Data Logger (Start using Button Press)
7	.10.	Configuration of a Data Logger (Real Time Logging)
8.	Μοι	unting Guidelines
9.	Bat	tery Replacement
10.	Т	echnical Support

#### 1. Document Overview

This guide is provided to distribute the following information:

- the main features of data logger
- how to install the software for the logger
- how to use the software for configuration, data extraction and data viewing
- how to mount your data logger
- how to replace the battery for the data logger
- contact information should you have further technical questions

#### 2. Product Overview

The OM-90 Series are compact temperature / humidity data loggers measuring just 35mm x 60mm x 15mm. These data loggers are commonly used in transport applications including but not limited to transport of:

- sensitive/fragile items
- food
- animal/livestock
- plants
- chemicals
- medicines or organs

The OM-90 Series comes in two versions; OM-91 is a temperature only version and OM-92 is a temperature and humidity logger. This document applies for both versions.



Figure 1: Introducing the OM-90 Series Data Loggers and software

Suppose for example you would like to transport some delicate or fragile cargo from one location to another. By attachment of a data logger one can record temperature and humidity conditions during the journey. At the destination you can then extract the logged data using very simple software and see graphically or in raw data format the changes in environment over time.

When you would like to guarantee that items are transported within particular environmental limits then one can also set operating ranges above or below which light indicators will be activated so that the user can become aware that the desired transportation conditions have been exceeded.



Figure 2: Data Logger Indicators

The data logger offers accurate and repeatable logging for temperature and relative humidity. Temperature is measured to an accuracy of  $\pm 0.3^{\circ}$ C over the range  $+5^{\circ}$ C to  $+60^{\circ}$ C. Over the full range  $-30^{\circ}$ C to  $+80^{\circ}$ C the accuracy is  $\pm 2.0^{\circ}$ C. Relative humidity is measured to an accuracy of  $\pm 2.0^{\circ}$  over 20% to 80% humidity range and  $\pm 3.0^{\circ}$  over 0% to 100%. Both temperature and relative humidity are logged at a user configurable logging rate which can be set in software.

The data logger is equipped with the largest logging memory when compared with comparable portable logging devices currently available. The data logger has provision for an impressive 65520 records, each of which is a time-stamped recording of temperature and humidity.

#### 3. Technical Specifications

•	Temperature range Temperature accuracy (limited range)	-30°C to +80°C (-22° F to 176° F) <u>+</u> 0.3°C (+5°C to +60°C) +0.5° F (41°F to 140°F)
•	Temperature accuracy (-30°C to +80°C) Temperature resolution	see accuracy curve in figure 3 0.01°C (0.2° F)
• • •	Humidity range (only on OM-92) Relative humidity accuracy (limited range) Relative humidity accuracy (0% to 100%) Relative humidity resolution	0% to 100% <u>+</u> 2.0% (20% to 80%) see accuracy curve in figure 3 0.01%
•	Temperature alarms (minimum & maximum) Relative humidity alarms (minimum & maximum)	
•	Battery Battery life	CR2540 coin cell up to 4 years while logging
•	Physical size	35mm x 60mm x 15mm
•	Software available on Windows XP, Vista, Windows 7, 8 & 10 (coming soon to Linux & MacOs)	
•	Logger identifier	16 characters
•	Logging space (includes time plus temperature & humidity)	65520 points





Humidity Accuracy (at 25°C)



Figure 3: Temperature and Humidity Accuracy

#### 4. Buttons and LED Indicators

The data logger includes one button located on the side of the logger. It is possible by configuration in the software to use this button as a way to start and stop logging. See section 7.9 Configuration of a Data Logger (Start using Button Press)



Figure 4: Data Logger Start/Stop Button

The logger contains two LED indicators - LED left and LED right with the following functionality:

#### <u>Status LED</u>

- When the logger starts logging then this LED pulses green once for 1 second.
- When logging this LED pulses green for 7 milliseconds and off for 6 seconds.
- When the logger stops then this LED pulses red for 1 second then off forever.

#### <u>Alarm LED</u>

• When a temperature or humidity alarm is triggered this LED pulses **red** for 7 milliseconds and off for 6 seconds.



Figure 5: Data Logger LED indicators

#### 5. Packaging List

Your data logger comes standard with the following items.

• One Data Logger unit (either OM-91 or OM-92)



Figure 6: Data Logger Unit

• One CR2450 coin cell battery

One installation CD (Software is also available for download on the OM-90 web page)





Figure 7: CR2450 Coin Cell Battery and Install CD

• One USB cable – type A to micro-USB type B



Figure 8: USB Cable

#### 6. Installation of the Software

Installation of the software is very simple. The data logger needs no special drivers since it is a USB HID device and such devices can make use of the standard drivers that come with Windows. Simply install the software from the CD included with the data logger. OR download and install the software from Omega's OM-90 web page. (look for the software icon).

In order to use this software consult the next section – Section 7 "Using the Software".

#### 7. Using the Software

#### 7.1. Front Page

When the software is started the welcome screen appears and gives a brief introduction to the data loggers and what they can be used for.

- Click on the Configuration Wizard Button to start configuration of the logger.
- Click on the Graph Button if you want to view data graphically.
- Click on the Raw Data Button if you want to view logger data as a table.



Figure 9: OM-90 Series Data Logger Software

#### 7.2. Main Page

The main page of the software consists of a configuration pane, a graphical pane and a spreadsheet pane. It is possible to have any of these enabled or disabled by the buttons located top left in the window. The panes will automatically start in an enabled or disabled view depending on the viewing mode select from the front page.



Figure 10: Software Main Page

Buttons are also included for saving/loading configuration and logged data to and from file.

There is also a button for selection between Celsius or Fahrenheit temperature scales.

A connection indication on the right hand side becomes blue when your application is connected to a data logger and is greyed-out otherwise.

Finally a home icon provides a link back to the front page.

#### 7.3. Configuration of a Data Logger (Start on Disconnection)

Configuration of the data logger starts by specifying the device name. Up to 16 characters is allowed in the name.

Logger         OM-92 (Temperature & Humidity)           Device ID         1135           PC Time         2014.11.05 13:06:09           Logger Time         2014.11.05 13:06:08	Temp [C]		Humidity [%]	DateTime Temp [C] Humidity
Logging Start Button Press   Logging Interval 1 sec (max 18 hours)   Temperature	83.33		83.33	
Enable  Alarm Threshold (High) disabled Alarm Threshold (Low) disabled Used	66.67-			
Humuly Enable I Alarm Threshold (High) 85 % Alarm Threshold (Low) 5 %	50.00		50.00	
Status Firmware Revision 20005 Battery Level 100%				
	33.33-			
Configure Logger Settings	15.67			
Read Logger Configuration/Data	0.00	30 40 50 60 70	0.00	

Figure 11: Configuration Step (Start on Disconnection)

Next select the logger start to On Disconnection. There are other options (At Specified Time, Button Press and Real Time Logging) which will be discussed in later sections of this manual.

Select the logging interval (this is the time between samples).

For temperature logging then one needs to enable or disable the logging. In the case the temperature logging is enabled then one can also set alarm thresholds for which the alarm indication will signify that temperature limits have been exceeded.

Similarly for humidity one needs to enable or disable the logging and configure the alarm limits.

			.ogger Busy
gger OM-92 (Temperature & Humidity) /ce ID 1135 Time 2014.11.05 13:08:10	Temp [C] Temperature / Humidity	Humidity [%]	DateTime Temp [C] Humidity
gger Ime 2014.11.05 13:08:09 gging Start Button Press ▼ gging Interval 1 sec (max 18 hours) ▼ nperature	83.33	63.33	
Alarm Threshold (High) disabled Alarm Threshold (Low) disabled anidity	66.67 Configuration	66.67	
Alarm Threshold (High) 85 %	50.00-	50.00	
tery Level	33.33-	-93,33	
Configure Logger Settings	15, 67-	-15.67	
Read Logger Configuration/Data	0.00 0 10 20 50 40 50 60		

Finally to save the configuration setup to the data logger, press the Configure Button.

Figure 12: Write Configuration to the Logger

Once configured the software will show the logger time will be synchronized to the PC time. The logger firmware revision will also appear in the application.

Once you disconnect the logger from the PC, logging will start automatically and at the configured logging interval. A green light will flash every 6 seconds to indicate that the unit is logging. The red light will flash also in the case that the temperature or humidity goes outside of the allowed limits and triggers the alarms.

If the logger becomes full then all lights will stop flashing.

#### 7.4. Saving a Configuration File

Configurations can be saved to file so that you can use the configuration next time or if you have multiple devices with the same configuration.

Saving a configuration file is just a matter of selecting the save icon and filling out the file name. If the logger contains no logged data then it will only save the configuration settings. Configuration and logged data is stored in Comma Separated Value (CSV) format.



Figure 13: Saving a Configuration to File

#### 7.5. Loading a Configuration from File

Loading a configuration file from disk is just a matter of selecting the open icon and selecting the configuration file name. This upload presents a simple way to configure multiple loggers with the same configuration. This ensures a common configuration and saves time/effort when configuring multiple devices.

ogger         OM-92 (Temperature & Humidity)           Device ID         1135           PC Time         2014.11.05 21:05:59           ogger Time         2014.11.05 21:05:58	Temp [C]	Temperature / Humidity	Hur	Da iidity [%] 100.00	teTime	Tem
ogging Start Button Press	Load File	rer ► Windows7_OS (C:) ► Data	▼ 4 <sub>2</sub>	Search Data		×
Enable V Alarm Threshold (High) disat Alarm Threshold (Low) disat Iumidity Enable V Alarm Threshold (High) 85 9 Alarm Threshold (Low) 5 % Itatus	Crganze Vew fold	Name All 1135_config.csv		Bes Date modified 11/5/2014 9:04 PM 11/5/2014 2:42 PM	Type Micros Micros	oft Exce
Battery Level 100%	Subversion Videos Homegroup Computer Kundows7_OS (C DVD RW Drive (D =					•

Figure 14: Loading a Configuration from File

Remember to press the configuration button if you want the configuration data held in the software to be written to your data logger.

#### 7.6. Retrieving Logged Data

Data can be retrieved from the logger by connecting to the logger and then pressing the read button at the bottom left of the software.

Both the device configuration and logged data will immediately become uploaded and viewable in the configuration, graphical and spreadsheet views.

Zooming in and out of the graphs can be achieved by the following controls:

- To Zoom In Drag the left mouse button
- To Zoom Out Click the right mouse button
- When zoomed in you can move around using the key-board arrows or the mouse scroll wheel

To write data and configuration to file, press the save icon. Data and configuration is then stored in Comma Separated Value (CSV) format which can be edited by third party software such as Microsoft Excel or read back into the data logger software at a later point.

ogger OM-92 (* evice ID C Time 2014.1	Temperature & Humidity) 1135 1.05 14:41:08	Temp [C] 31.75	Temperature / Humidity	Humid	ty [%] 1 100,00 2 3	DateTime 2014-06-15T10:48:20 2014-06-15T10:48:30 2014-06-15T10:48:40
ogger Time 2014.1	1.05 14:41:08				X 5	2014-06-15T10:49:00 2014-06-15T10:49:10
ogging Start Buttor				1	6	2014-06-15T10:49:20
ogging Interval 1 sec (	🕻 🌍 🔵 🗢 📕 🕨 Computer 🕨 Win	dows7_OS (C:) ► Data	✓ ✓ Search Data		P 7	2014-06-15T10:49:40
amparatura	Organize  New folder			8== +	0 8	2014-06-15T10:49:50
emperature	Be	*			9	2014-06-15T10:50:00
nable 🔽	A Documents A Name		Date modified	Туре	10	2014-06-15T10:50:10
	A Music	Ma itania in	abab states and an		11	2014-06-15T10:50:20
	Pictures	NO LEMS M	aten your searen.		12	2014-06-15T10:50:30
umidity	Subversion				13	2014-06-15T10:50:40
nable 🔽	A Videos				14	2014-06-15T10:50:50
					15	2014-06-15T10:51:00
	1 😽 Homegroup				16	2014-06-15T10:51:10
atus					17	2014-06-15T10:51:20
	Computer				18	2014-06-15T10:51:30
rmware Revision	Windows7_OS (C				19	2014-06-15T10:51:40
attery Level	DVD RW Drive (D				20	2014-06-15T10:51:50
	B Lenovo_Recovery				21	2014-06-15T10:52:00
					22	2014-06-15110:52:10
	📬 Network				23	2014-06-15110:52:20
		m	1		1 24	2014-06-15110:52:30
					25	2014-00-13110:52:40
_	File name: 1135_data				▼ 20 27	2014-06-15110:52:50
Configure Logar	Save as type: CSV(*.csv)			_	- 2/	2014-06-15110:53:00
Settings					20	2014-06-15T10:53:10
	Hide Folders		Save	Cancel	30	2014-06-15T10:53:20
				-	50	2014-00-10110.00.00

Figure 15: Saving Logged Data to File

#### 7.7. Viewing Logged Data from File

It is a simple matter to read back data from file (CSV format) into the software and view previous logs graphically or using the spreadsheet views. Simply press the open icon and then select the file which you would like to view.



Figure 16: Retrieving Logged Data from File

Omega OM-90 Series Data Logger Software (1.0.16)			
		Ð	1
Ingger OM-02 (Temperature & Humidity)	DateTime	Temp [C]	Humidity
Temp [C]	Humidity [%] 1. 2014-06-15T10:48:20	24.98	38.98
31.75	100.00 2 2014-06-15T10:48:30	25	39.19
C Time 2014.11.05 12:45:22	3 2014-06-15T10:48:40	25.01	38.7
ogger Time 2014.11.05 12:45:22	4 2014-06-15T10:49:00	25.01	38.8
naging Start Button Press	5 2014-06-15T10:49:10	25.02	38.34
	6 2014-06-15T10:49:20	25.02	38.17
ogging Interval 1 sec (max 18 hours)	-33.33 7 2014-06-15T10:49:40	25	37.82
magniture .	8 2014-06-15T10:49:50	24.99	39.4
rature	9 2014-06-15T10:50:00	24.99	38.53
nable 🗹 🛛 Alarm Threshold (High) disabled 🔹	10 2014-06-15T10:50:10	24.99	38.2
Alarm Threshold (Low) disabled	11 2014-06-15T10:50:20	24.99	38.04
Alarm Threshold (Low) disabled  21.15	-66.67 12 2014-06-15T10:50:30	24.97	37.9
	13 2014-06-15T10:50:40	24.95	37.74
able 🖉 Alarm Threshold (High) 85 %	14 2014-06-15T10:50:50	24.95	37.7
	15 2014-06-15T10:51:00	24.94	37.58
Alarm Threshold (Low) 5 %	16 2014-06-15T10:51:10	24.93	37.59
15.87-	-50.00 17 2014-06-15T10:51:20	24.93	37.53
	18 2014-06-15T10:51:30	24.93	37.40
mware Revision 20005	19 2014-06-15T10:51:40	24.92	37.49
ttery Level 100%	20 2014-06-15T10:51:50	24.92	37.4
	21 2014-06-15T10:52:00	24.91	37.39
10.58-	-33.33 22 2014-06-15T10:52:10	24.91	37.35
	23 2014-06-15T10:52:20	24.9	37.33
	24 2014-06-15T10:52:30	24.9	37.3
	25 2014-06-15T10:52:40	24.89	37.34
	26 2014-06-15T10:52:50	24.88	37.30
5.29	-16.67 27 2014-06-15T10:53:00	24.88	37.34
Configure Logger	28 2014-06-15T10:53:10	24.88	37.35
Settings	29 2014-06-15T10:53:20	24.87	37.34
	30 2014-06-15T10:53:30	24.87	37.3
Read Logger	31 2014-06-15T10:53:40	24.86	37.2
Configuration/Data	0.014 15 0901	24.86	37.20
100%, 10:43:20 11:53:10	12:58:00 14:02:40 15:07:20 16:12:30 17:17:10 18:21:50 19:26:30 20:31:10 33 2014-06-15T10:54:00	24.86	37.29
Temperatu	re Humidity 34 2014-06-15T10:54:10	24.85	37.24

Figure 17: Retrieving Logged Data from File (complete)

#### 7.8. Configuration of a Data Logger (Start at Specified Time)

You can configure a device so that it starts at a specific time. Configure as per section 7.3 but select At Specified Time. Fill in the exact month, day, hour, minute and seconds when you want the logging to start.

Remember to write the setup to the logger by pressing the Configure Button.

When you have disconnected the logger and the time reaches the desired start time for the logging then the logger will begin to log. This is a perfect way to have multiple loggers start logging at exactly the same time.

					1
Logger         OM-92 (Temperature & Humidity)           Device ID         1135           PC Time         2014.11.05 14:46:14           Logger Time         2014.11.05 14:46:13	Temp [C] 100.00	Temperature / Humidity	Humidity (%)	DateTime	Temp
.ogging Start At Specified Time ▼ 11.15 14:50:00 🔄 .ogging Interval 10 secs (max 7 days) ▼ Femperature	83.33		83,33		
Enable V Alarm Threshold (High) disabled  Alarm Threshold (Low) disabled	66.67				
Enable 🗹 Alarm Threshold (High) 85 % 🔹 Alarm Threshold (Low) 5 %	50.00				
Itatus Firmware Revision 20005 Battery Level 100%					
Configure Logger Settings					
Read Logger Configuration/Data	0.00	20 30 40 50 60 70 8	0.00		

Figure 18: Configuration Step (Start at Specified Time)

#### 7.9. Configuration of a Data Logger (Start using Button Press)

You can also configure the device so that it starts logging on a button press. Configure as per section 7.3 but select Button Press.

Remember to write the setup to the logger by pressing the Configure Button.

When you have disconnected the logger you can start logging by a simple button press. You will then see it is logging because the green light starts flashing every 6 seconds. You can also stop it using a button press – and all lights will stop flashing.

ogger OM-92 (Temperature & Humidity)	1000	Temperature / Humidity		DateTime	Temp
evice ID 1135	Temp [C] 100.00		Humidity [%]		
Time 2014.11.05 14:46:55			and the second		
gger Time 2014.11.05 14:46:54					
gging Start Button Press					
gging Interval 10 secs (max 7 days) 🔻	83.33		83.33		
mperature					
able 🗹 Alarm Threshold (High) disabled 🔻					
Alarm Threshold (Low) disabled	65.67				
midity					
able 🗹 🛛 Alarm Threshold (High) 😣 🗸					
Alarm Threshold (Low) 5 %					
atus	50.00				
mware Revision 20005					
ttery Level 100%					
	33.33				
	and the second s				
	16.67				
Settings					
North Production	100 March 100				
Configuration/Data	10.00		0.00		

Figure 19: Configuration Step (Start using Button Press)

#### 7.10. Configuration of a Data Logger (Real Time Logging)

You may also use your data logger in Real Time Logging mode. Select the start logging to Real Time Logging with the appropriate logging interval. Next press start. You will then see the software take samples from the logger directly and display them on the table and graphs. This is a way to monitor temperature and humidity conditions directly from your PC instead of using the remote logging.



Figure 20: Configuration Step (Real Time Logging)

#### 8. Mounting Guidelines

There is no particular limitation with regard to transporting your data logger. The logger may be fastened to a transportation container or loosely packed.

It is recommended however to keep some air space around the ventilation slots on the logger or else the logger may take longer to track humidity and temperature changes.

#### 9. Battery Replacement

Your data logger uses a CR2450 3V Lithium-Ion battery.

In order to replace the battery:

- First un-mount the logger.
- Next the cover needs to unscrewed and removed.
- The old battery should then be removed and a new battery inserted. Please try not to touch any exposed electronics in the process.
- Finally attach the cover again with screws.

# WARRANTY/DISCLAIMER

OMEGA ENGINEERING, INC. warrants this unit to be free of defects in materials and workmanship for a period of **13 months** from date of purchase. OMEGA's WARRANTY adds an additional one (1) month grace period to the normal **one (1) year product warranty** to cover handling and shipping time. This ensures that OMEGA's customers receive maximum coverage on each product.

If the unit malfunctions, it must be returned to the factory for evaluation. OMEGA's Customer Service Department will issue an Authorized Return (AR) number immediately upon phone or written request. Upon examination by OMEGA, if the unit is found to be defective, it will be repaired or replaced at no charge. OMEGA's WARRANTY does not apply to defects resulting from any action of the purchaser, including but not limited to mishandling, improper interfacing, operation outside of design limits, improper repair, or unauthorized modification. This WARRANTY is VOID if the unit shows evidence of having been tampered with or shows evidence of having been damaged as a result of excessive corrosion; or current, heat, moisture or vibration; improper specification; misapplication; misuse or other operating conditions outside of OMEGA's control. Components in which wear is not warranted, include but are not limited to contact points, fuses, and triacs.

OMEGA is pleased to offer suggestions on the use of its various products. However, OMEGA neither assumes responsibility for any omissions or errors nor assumes liability for any damages that result from the use of its products in accordance with information provided by OMEGA, either verbal or written. OMEGA warrants only that the parts manufactured by the company will be as specified and free of defects. OMEGA MAKES NO OTHER WARRANTIES OR REPRESENTATIONS OF ANY KIND WHATSOEVER, EXPRESSED OR IMPLIED, EXCEPT THAT OF TITLE, AND ALL IMPLIED WARRANTIES INCLUDING ANY WARRANTY OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE HEREBY DISCLAIMED. LIMITATION OF LIABILITY: The remedies of purchaser set forth herein are exclusive, and the total liability of OMEGA with respect to this order, whether based on contract, warranty, negligence, indemnification, strict liability or otherwise, shall not exceed the purchase price of the component upon which liability is based. In no event shall OMEGA be liable for consequential, incidental or special damages.

CONDITIONS: Equipment sold by OMEGA is not intended to be used, nor shall it be used: (1) as a "Basic Component" under 10 CFR 21 (NRC), used in or with any nuclear installation or activity; or (2) in medical applications or used on humans. Should any Product(s) be used in or with any nuclear installation or activity, medical application, used on humans, or misused in any way, OMEGA assumes no responsibility as set forth in our basic WARRANTY/DISCLAIMER language, and, additionally, purchaser will indemnify OMEGA and hold OMEGA harmless from any liability or damage whatsoever arising out of the use of the Product(s) in such a manner.

# **RETURN REQUESTS/INQUIRIES**

Direct all warranty and repair requests/inquiries to the OMEGA Customer Service Department. BEFORE RETURNING ANY PRODUCT(S) TO OMEGA, PURCHASER MUST OBTAIN AN AUTHORIZED RETURN (AR) NUMBER FROM OMEGA'S CUSTOMER SERVICE DEPARTMENT (IN ORDER TO AVOID PROCESSING DELAYS). The assigned AR number should then be marked on the outside of the return package and on any correspondence.

The purchaser is responsible for shipping charges, freight, insurance and proper packaging to prevent breakage in transit.

FOR **WARRANTY** RETURNS, please have the following information available BEFORE contacting OMEGA:

- 1. Purchase Order number under which the product was PURCHASED,
- 2. Model and serial number of the product under warranty, and
- 3. Repair instructions and/or specific problems relative to the product.

FOR **NON-WARRANTY** REPAIRS, consult OMEGA for current repair charges. Have the following information available BEFORE contacting OMEGA:

- 1. Purchase Order number to cover the COST of the repair,
- 2. Model and serial number of the product, and
- 3. Repair instructions and/or specific problems relative to the product.

OMEGA's policy is to make running changes, not model changes, whenever an improvement is possible. This affords our customers the latest in technology and engineering.

OMEGA is a registered trademark of OMEGA ENGINEERING, INC.

© Copyright 2017 OMEGA ENGINEERING, INC. All rights reserved. This document may not be copied, photocopied, reproduced, translated, or reduced to any electronic medium or machine-readable form, in whole or in part, without the prior written consent of OMEGA ENGINEERING, INC.

# Where Do I Find Everything I Need for Process Measurement and Control? OMEGA...Of Course! Shop online at omega.com

## TEMPERATURE

- Thermocouple, RTD & Thermistor Probes, Connectors, Panels & Assemblies
- Wire: Thermocouple, RTD & Thermistor
- Calibrators & Ice Point References
- Recorders, Controllers & Process Monitors
- Infrared Pyrometers

## **PRESSURE, STRAIN AND FORCE**

- Transducers & Strain Gages
- ☑ Load Cells & Pressure Gages
- Displacement Transducers
- Instrumentation & Accessories

## **FLOW/LEVEL**

- Rotameters, Gas Mass Flowmeters & Flow Computers
- Air Velocity Indicators
- Turbine/Paddlewheel Systems
- Totalizers & Batch Controllers

# pH/CONDUCTIVITY

- 🗹 pH Electrodes, Testers & Accessories
- Benchtop/Laboratory Meters
- Controllers, Calibrators, Simulators & Pumps
- 🗹 Industrial pH & Conductivity Equipment

# DATA ACQUISITION

- Communications-Based Acquisition Systems
- Data Logging Systems
- Wireless Sensors, Transmitters, & Receivers
- Signal Conditioners
- Data Acquisition Software

#### HEATERS

- Heating Cable
- Cartridge & Strip Heaters
- Immersion & Band Heaters
- Flexible Heaters
- Laboratory Heaters

## ENVIRONMENTAL MONITORING AND CONTROL

- Metering & Control Instrumentation
- Refractometers
- Pumps & Tubing
- Air, Soil & Water Monitors
- 🗹 Industrial Water & Wastewater Treatment
- PH, Conductivity & Dissolved Oxygen Instruments